

\*\*\*\*\* **CERTIFICATE OF ANALYSIS** \*\*\*\*\*

Article:	CompoFlow® 4F T&B 63 ml CPD/100 ml SAGM - RCC - PDS-V		
Article number:	CQ32250		
Anticoagulant solution:	63 ml CPD		
Preservation solution:	100 ml SAG-Mannitol		
Batch number:	85WF04FE00	Quantity:	13816
Manufacturing date:	2024.06	Expiry date:	2026.05

Test results

1. CPD Anticoagulant solution

	Method <sup>1</sup>	Nominal value	Limits	Results
Sodium citrate.2H <sub>2</sub> O	Titration	26.3 g/l	25.0 - 27.6	25.6 g/l
Citric acid.1H <sub>2</sub> O	Titration	3.27 g/l	2.94 - 3.60	3.22 g/l
NaH <sub>2</sub> PO <sub>4</sub> .2H <sub>2</sub> O	UV	2.51 g/l	2.38 - 2.64	2.44 g/l
Glucose.1H <sub>2</sub> O <sup>2</sup>	HPLC	25.5 g/l	24.2 - 26.8	24.8 g/l
5-HMF	HPLC		< 5 ppm	0.4 ppm
pH	Potent	5.6	5.3 - 5.9	5.7
Volume	Volumetr.	63 ml	57 - 69	67 ml

2. SAG-Mannitol Preservation solution

	Method	Nominal value	Limits	Results
Sodium chloride	Argent.	8.77 g/l	8.33 - 9.21	8.66 g/l
Adenine	UV	0.169 g/l	0.161-0.177	0.165 g/l
Glucose.1H <sub>2</sub> O <sup>2</sup>	HPLC	9.00 g/l	8.55 - 9.45	8.80 g/l
Mannitol	HPLC	5.25 g/l	4.99 - 5.51	5.12 g/l
5-HMF	HPLC		< 5 ppm	0.3 ppm
pH	Potent.	5	3.5 - 6.0	5.2
Volume	Volumetr.	100 ml	90 - 110	106 ml

<sup>1</sup>) The used test methods are equivalent or more accurate than the corresponding methods described in the Eur. Pharm.

<sup>2</sup>) As total reducing sugars

- |  |                    |
|--|--------------------|
| 3. Characteristics: clear, colourless to faintly yellow solution practically free from visible particles | conform            |
| 4. Physical control: in accordance with specifications   | conform            |
| 5. Sterility (Eur. Pharm.): sterile  | conform            |
| 6. Bacterial Endotoxines (Eur. Pharm.): CPD: < 5.56 EU/ml<br>SAGM: < 5.56 EU/ml                          | conform<br>conform |
| 7. Used Filterbatch  | 24052104           |

<u>Batch released by Quality Assurance.</u>	<u>Manufactured by:</u>
Date: 03 July 2024	Fresenius Kabi AG
Name: Xu Xiaoying	Plant: FKGZ
Signature: 	