



**FRESENIUS  
KABI**

\*\*\*\*\* **CERTIFICATE OF ANALYSIS** \*\*\*\*\*

Article: CompoStop Flex2F PLT processing and leuko red.syst  
 Article number: FD51600  
 Batch number: 41WI17GA00 Quantity: 14 320  
 Manufacturing date: 27 September 2024 Expiry date: 2026-08

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	24PA015	

<u>Batch released by Quality Assurance.</u>	<u>Manufactured by:</u>
Date: 18.10.2024	Fresenius Kabi AG
Name: Tykva František	Plant: Fresenius Kabi Horatev CZ s.r.o.
Signature: 	Hořátev 104, 289 13
	Czech Republic