

\*\*\*\* **CERTIFICATE OF ANALYSIS** \*\*\*\*

Article:	CompoFlex® 1F Empty - PD-V		
Article number:	T2131		
Batch number:	85WG11CA00	Quantity:	5558
Manufacturing date:	2024.07	Expiry date:	2027.06

<b>Test</b>	<b>Limits</b>	<b>Results</b>
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 31 July 2024 Name: Xu Xiaoying Signature: 	<u>Manufactured by:</u> Fresenius Kabi AG Plant: FKGZ
--	---